



Fiscal Year 2020

CERTIFICATION OF REGISTRATION

This certifies that:

Name: Wuchuan qinbawehui Labor Protection Appliance Factory

Add: No.155, Qincun, Qinba Town, Wuchuan City

has completed the FDA Establishment Registration (as manufacturer and foreign exporter) and Device Listing with the US Food & Drug Administration, through

The Owner/ Operator Number for this Registration is : 10064648

Listing No	Code	Device Name	Product Name
D379589	MSH	Respirator, surgical	8200 high efficiency dust-proof cup mask Disposable high-efficiency particle respirator Self priming filter type anti particle respirator

ABmed will confirm that such registration remains effective upon request and presentation of this certificate until the end of the year stated above, unless said registration is terminated after issuance of this certificate. ABmed makes no other representations or warranties, nor does this certificate make any representations or warranties to any person or entity other than the named certificate holder, for whose sole benefit it is issued. This certificate does not denote endorsement or approval of the certificate - holder' s device or establishment by the U.S Food and Drug Administration.

ABmed assumes no liability to any person or entity in connection with foregoing.

Date of verification:Mar. 27, 2020

Date of expiration:Dec. 31, 2020

SH OFFICE

TEL:0086-21-50313932 Boyle Wang Phone:0086-18930777676 info@truthful.com.cn

ABMED SERVICE INC.

36 Soyth 18th Avenue, Suite A Brighton,CO USA 80601

TEL:213-375-3998 FAX:213-375-3998 info@abmed.com.cn



Technical Construction File no. TPPEMJ2003123551-AG02-28

Wuchuan Qinbawehui Labor
Protection Appliance Factory
No.155, Qincun, Qinba Town, Wuchuan City

Self priming filter type anti particle respirator
9801

Standard:
EN 149:2001+A1:2009

related to CE Directive(s):
R 2016/425 (Personal Protective Equipment)

Additional information and clarification about the Marking:



Expiry date: 13 March 2025





Filter material: KN95 protective level



Wei Hui high-efficiency dust-proof folding mask

Effectively resist PM2.5 dust haze and so on

Good quality Wei Huizao

2 pack
pcs

For the sake of your health, we recommend that you have a simple 4-part test before use.

1. The overall appearance of the mask is not damaged or seriously contaminated;
2. The headband is undamaged and elastic;
3. The nose clip part is not broken or broken;
4. Check the status of Suction Valve and valve plate if there is one.

Proper way to wear the ear mask, please follow the steps below

1. Clamp the nose outwards, holding one ear band in each hand to make sure the nose clamp is upward;
2. Put the mask on, place the Chin under the mask, and place the ear straps behind the ears with both hands;
3. Adjust to comfortable position, make the mask close to the face;
4. Press the index and middle fingers of both hands to adjust the nose clip, until they are close to the bridge of the nose;
5. For Air tightness testing.

Scope of use and storage

1. For adults;
2. Do not wash;
3. Store in a well-ventilated, dark and dry environment, away from fire and pollution;
4. RECOMMENDED STORAGE CONDITIONS: (temperature: -10℃ ~ 40℃ relative humidity less than 85%);
5. Shelf Life: 3 years (production date see certificate of conformity)
6. This product conforms to the National Standard: GB2626-2006 KN95

Manufacturer: Wuchuan City Qinda Weihui Labor Protection Products Factory
Production address: No. 155, Qinda Village, Qinda Town, Wuchuan City, Guangdong Province
Client: Foshan Nanhai Weihui Labor Protection Products Co., Ltd.
Tel: 0757-81826778 81826782

* Rank: KN95



6 972697 400446

CE FDA



Warning

Masks can filter out certain contaminants, But it doesn't eliminate the risk of disease or infection. Incorrect use of mask, Can Lead to disease and even death. Materials in direct contact with the skin may cause allergic reactions in some sensitive individuals.



Foshan Weihui Labor Protection Products Co., Ltd.

伟辉



Up



Fear of
getting wet



Fragile



Accumulation
limit



Trample
protection



Product name:	High-efficiency foldable particle-proof respirator
model:	9801
Quantity:	500x(50x10box)
Volume:	680x305x315mm
Manufacturer:	Wuchuan City Qinba Weihui Labor Protection Products Factory
CONSIGNOR:	Client: Foshan Nanhai Weihui Labor Protection Products CO., Ltd.
TEL:	0757-818126778
Production address:	No.155, Qinba Village, Qinba town, Wuchuan City, Guangdong Province
National recruitment agents	

It's greener ● More durable ● Healthier



®

9801

Standard size/Earloop
Particulate Respirator

伟辉

Self-priming filter anti-particle inhaler
Disposable face mask (mask)
non-breathing Valve

GB2626-2006 KN95



Standard size / Earloop

Particulate Respirator


50 PCS

The National Quality Supervision and Test Center for Personal Protection Equipment(Beijing)

Test Report

Test Report No.: No.2020 (Dust)-541 (EN)

Page 2 of 2

Page 2 of 2									
No.	Test Item	Standard Clause	Requirement	Test Result			Conclusion	Note	
1	Filter efficiency	5.3	KN90≥ 90.0% KN95≥ 95.0% KN100≥ 99.97% Flow conditioning: Single filter: (85±4) L/min	Sample number		Initial filter efficiency (η)	Loading filter efficiency (η)	——	——
						1#	97.9 %		
				2#	97.8 %	95.2 %			
				3#	97.9 %	95.7 %			
				4#	97.7 %	96.8 %			
				5#	97.8 %	96.9 %			
				6#	97.8 %	95.8 %			
				7#	97.7 %	96.8 %			
				8#	97.8 %	95.5 %			
				9#	97.8 %	95.8 %			
				10#	98.0 %	96.8 %			
					KN: Ambient temperature: (25±5)℃ Relative humidity: (30±10) %	Ambient temperature: 24 ℃ Relative humidity: 32 %			
2	Inhalation resistance	5.5	≤350Pa	As received	3#	76 Pa	——	——	
					4#	76 Pa			
3	Exhalation resistance	5.5	≤250Pa	As received	3#	35 Pa	——	——	
					4#	41 Pa			
Sample photo									
Note			——						


Test Date: From 2020.02.22 to 2020.02.23

End

检验报告专用章

CAUTION

1. THE REPORT IS INVALID WITHOUT TEST REPORT SEAL AND SEAL OF THE TEST AGENCY.
2. ANY COPY OF THE REPORT IS INVALID WITHOUT TEST REPORT SEAL AND SEAL OF THE TEST AGENCY.
3. THE REPORT IS INVALID WITHOUT SIGNATURES OF MAIN INSPECTOR, AUDITER AND THE TEST APPROVAL.
4. THE TEST REPORT IS INVALID IF MODIFIED.
5. APPLICANT TEST REPORTS REFER ONLY TO THE SAMPLES TESTED.
6. PARTIAL COPY OF THE REPORT IS NOT ALLOWED WITHOUT WRITTEN AUTHORIZATION OF THE TEST AGENCY.



Account Name: The PRC Quality Supervision and Testing Center for Personal Protection Equipment

Bank: ICBC Beijing Taoranting Branch

Account: 0200049109024901958

Add: No. 55 Taoranting Road, Xicheng District, Beijing, China

Post Code: 100054

Tel: +86-10-63520770

Fax: +86-10-63520770

Email: bjlb63520770@VIP.163.com



160010260248



(2016)国认监认字(099)号



中国认可
国际互认
检测
TESTING
CNAS L1499

Test Report

Test Report No.: No.2020(Dust)-541(EN)

PRODUCT: Respiratory Protective Equipment-Non-Powered
Air-Purifying Particle

APPLICANT: Wuchuan, Qinba Weihui Labor Protection
Appliance Factory

TEST TYPE: Applicant Test

THE NATIONAL QUALITY SUPERVISION
AND TEST CENTER FOR PERSONAL
PROTECTION EQUIPMENT(BEIJING)

(北京)

国家劳动保护用品质量监督检验中心（北京）
检 验 报 告



检验报告编号: No. 2020(生) - 541

共 2 页 第 1 页



样品名称	自吸过滤式防颗粒物呼吸器	商 标	伟辉
规格型号	随弃式面罩 KN95	防护级别	KN95
委托单位名称	吴川市覃巴伟辉劳保用品厂	委托单位电话	13822505678
委托单位地址	吴川市覃巴镇秦村 155 号		
生产单位名称	吴川市覃巴伟辉劳保用品厂	生产单位电话	13822505678
生产单位地址	吴川市覃巴镇秦村 155 号		
样品数量	20 个	生产日期	2020.02.11
样品状态	完好	到样日期	2020.02.22
样品特征	白色折叠口罩	送样者	黄伟辉
检验依据	GB 2626-2006《呼吸防护用品 自吸过滤式防颗粒物呼吸器》		
检验项目	过滤效率、吸气阻力、呼气阻力		
检 验 结 论	<p>该样品经检验,依据 GB 2626-2006《呼吸防护用品 自吸过滤式防颗粒物呼吸器》,判定所检项数据符合(KN95)标准要求。</p> <div style="text-align: center;">  <p>签发日期: 2020 年 02 月 23 日</p> </div>		
备 注	样品未做预处理		
<p>批 准: 庞平 审 核: 何强 主 检: 杨</p>			

国家劳动保护用品质量监督检验中心（北京）
检验报告（检验结果汇总）

检验报告编号：No. 2020(尘) - 541

共2页 第2页

共2页 第2页

序号	检验项目	标准条款号	标准要求	实 测 结 果			本项结论	备注	
1	过滤效率	5.3	KN90≥ 90.0% KN95≥ 95.0% KN100≥ 99.97% 测试气流量： 单过滤元件： (85±4) L/min	未做 预处理		初始过滤效率 (η)	加载过滤效率 (η)	符合 标准 要求 (KN95)	
					1#	97.9 %	96.8 %		
					2#	97.8 %	95.2 %		
					3#	97.9 %	95.7 %		
					4#	97.7 %	96.8 %		
					5#	97.8 %	96.9 %		
					6#	97.8 %	95.8 %		
					7#	97.7 %	96.8 %		
					8#	97.8 %	95.5 %		
					9#	97.8 %	95.8 %		
					10#	98.0 %	96.8 %		
			KN 类： 环境温度：(25±5)℃ 相对湿度：(30±10)%		环境温度： 24 ℃ 相对湿度： 32 %				
2	吸气阻力	5.5	每个样品的总吸气阻力 应≤350Pa。	未做 预处理	3#	76 Pa	符合 标准 要求		
				4#	76 Pa				
3	呼气阻力	5.5	每个样品的总呼气阻力 应≤250Pa。	未做 预处理	3#	35 Pa	符合 标准 要求		
				4#	41 Pa				
样品照片									
备 注									

检验日期：2020 年 02 月 22 日至 2020 年 02 月 23 日

结 束